

## Smart paper-based electrochemical (bio)sensors for sustainable detection of biomarkers

Fabiana Arduini<sup>1,2</sup>

<sup>1</sup>University of Rome Tor Vergata, Department of Chemical Science and Technologies, via della ricerca scientifica, Rome, Italy

<sup>2</sup>SENSE4MED, via Bitonto 139, Rome, Italy

[fabiana.arduini@uniroma2.it](mailto:fabiana.arduini@uniroma2.it)

---

As reported in my recent review entitled "Electrochemical paper-based devices: When the simple replacement of the support to print ecodesigned electrodes radically improves the features of the electrochemical devices" published in Current Opinion in Electrochemistry SI: Emerging Opinions (2022) [1]: *"Paper-based electrochemical (bio)sensors have emerged as highly attractive analytical devices for their superior sustainable features, such as avoiding the use of polyester as support and the reduction of waste, being incinerated after use. However, paper-based electrochemical (bio)sensors have recently demonstrated further advantages, including the simple combination with vertical microfluidics and their use as a reservoir to deliver smart electrochemical (bio)sensors able to i) contain the reagents, ii) preconcentrate the target analyte, and iii) synthesize the nanomaterials inside the paper network. Furthermore, these devices have demonstrated their ability to overcome the limitations of the other printed electrochemical sensors in the measurement of entirely liquid samples by detecting the target analyte in the aerosol phase or solid sample, without the additional sampling system. These achievements highlight their valuable and varied advantages in the sensing sector"*. During my presentation, I will report on the roadmap research activity carried out in the last 8 years related to the development of paper-based electrochemical devices as smart and sustainable point-of-care devices [2-6].

### Acknowledgements

PRIN2022 2022XTCTWH - SMARTMASK4CF

### References

[1] F. Arduini. Curr. Opin. Electrochem. 101090 (2022).

- [2] G. Scordo, D. Moscone, G. Palleschi, F. Arduini. Sens. Actuat. B, 258 (2018) 1015-1021.
- [3] S. Cinti, D. Moscone, F. Arduini. Nat. Protoc. 14.8 (2019) 2437-2451.
- [4] V. Caratelli, A. Ciampaglia, J. Guiducci, G. Sancesario, D. Moscone, F. Arduini. Biosens. Bioelectron. 165 (2020) 112411.
- [5] L. Fiore, V. Mazzaracchio, A. Serani, G. Fabiani, L. Fabiani, G. Volpe, D. Moscone, G.M. Bianco, C. Occhiuzzi, G. Marrocco, F. Arduini. Sens. Actuat. B 379 (2023) 133258.
- [6] L. Fiore, A. Sinha, N. Seddaoui, J. Biasio, F. Ricci, G.M. Stojanovic, F. Arduini. Chem. Comm. 59 (2023) 4300-4303.