

# Toward integrated detection and graphene-based removal of contaminants in a lab-on-a-chip platform

Andrzej Chałupniak<sup>1</sup>  
Arben Merkoçi<sup>1,2</sup>

arben.merkoci@icn2.cat  
andrzej.chalupniak@icn2.cat

<sup>1</sup> Catalan Institute of Nanoscience and Nanotechnology (ICN2)CSIC and The Barcelona Institute of Science and Technology, Campus UAB, Bellaterra, Barcelona, Spain


<sup>2</sup> ICREA, Pg. Lluís Companys 23BarcelonaSpain

In this work a miniaturized microfluidic platform for the simultaneous detection and removal of polybrominated diphenyl ethers (PBDEs) is developed. This device consists of a polydimethylsiloxane (PDMS) microfluidic chip for an immunoreaction step, a PDMS chip with an integrated screen-printed electrode (SPCE) for detection and a PDMS-reduced graphene oxide (rGO) chip for physical adsorption and subsequent removal of PBDE residues. The detection was based on competitive immunoassay (PBDE vs. HRP-PBDE) followed by the monitoring of enzymatic oxidation of o-aminophenol (o-AP) using square wave anodic stripping voltammetry (SW-ASV). PBDE was detected with the limit of detection similar to that obtained with a commercial colorimetric test (0.018 ppb), but with the advantage of using lower reagent volumes and a reduced analysis time. The use of microfluidic chips also provides improved linearity and a better reproducibility in comparison to those obtained with batch-based measurements using screen-printed electrodes. In order to design a detection system suitable for hazardous compounds such as PBDEs, a reduced graphene oxide–PDMS composite was developed and optimized to obtain increased adsorption (based on both the hydrophobicity and  $\pi$ – $\pi$  stacking between rGO and PBDE molecules) compared to those of non-modified PDMS. To the best of our knowledge, this is the first demonstration of electrochemical detection of flame retardants and a novel application of the rGO-PDMS composite in a biosensing system. This system can be easily adjusted to detect any analyte using the appropriate immunoassay and it supports operation in complex matrices such as seawater.


## References

[1] Chałupniak, A. & Merkoçi, A. Nano Res. (2017). doi:10.1007/s12274-016-1420-3

## Figures



**Figure 1.** Lab-on-a-chip platform for electrochemical detection and graphene-based removal of polybrominated diphenyl ether (PBDE).



**Figure 2.** rGO-PDMS composite (a) fabrication process, (b) a solid piece of rGO-PDMS, (c) rGO-PDMS-glass microfluidic chip